

**KENYA ELECTRICITY GENERATING COMPANY LIMITED**

**TENDER FOR SUPPLY OF SEISMOMETER EQUIPMENT,  
COMMISSIONING AND TRAINING FOR OLKARIA GEOTHERMAL  
DEVELOPMENT DIVISON – KGN-GDD-113-2017**

**4/12/2018**

**CLARIFICATION NO.2**

In accordance with the tender for supply of seismometer equipment, commissioning and training, KenGen hereby issues Clarification No. 2

	<b>CLARIFICATION SOUGHT</b>	<b>KENGEN'S RESPONSE</b>
1	Maximum frequency response bandwidth 0.0083 to 200hz (120s to 0.005s) As for the frequency, we could meet the requirement mentioned in the tender document, but based on the suggestion given by our professor, we recommend product with frequency 5-150Hz which has been sufficient for the operation, as product with frequency 0.0083 will be expensive and increase the cost, please advise whether our recommendation is acceptable?	The equipment should be able to meet the technical specifications ranges as outlined in the tender.
2	Operational temperature -45°C to +100°C or more As for the operational temperature, what we can meet is -45°C to + 85°C, whether is it acceptable?	

**ACKNOWLEDGEMENT OF CLARIFICATION NO.2**

We, the undersigned hereby certify that the addendum is an integral part of the document and the alterations set out in Addendum has been incorporated in the tender proposal.

Signed \_\_\_\_\_

Tenderer \_\_\_\_\_