



KENYA ELECTRICITY GENERATING COMPANY LIMITED

KGN-GDD-078-2018

TENDER FOR SUPPLY OF LIQUID DRILLING DETERGENT FOR GEOTHERMAL DEVELOPMENT DIVISION (TWO (2) - YEAR FRAMEWORK CONTRACT)

10th September 2018

CLARIFICATION NO.1 - VISCOSITY

In accordance with the TENDER FOR SUPPLY OF LIQUID DRILLING DETERGENT FOR GEOTHERMAL DEVELOPMENT DIVISION (TWO (2) - YEAR FRAMEWORK CONTRACT), KenGen hereby issues Clarification No.1

Bidder's Request	KenGen Response
Is Viscosity measured with Brookfield Viscometer?	Viscosity can be measured using a Kinematic or dynamic viscometer. The Model or Make of the Viscometer does not matter
At which rpm?	Calibrate the viscometer in use then select the rpms required
At which temperature?	Measured at room temperature

ACKNOWLEDGEMENT OF CLARIFICATION NO. 1

We, the undersigned hereby certify that the clarification is an integral part of the document and the alterations set out in the clarification has been incorporated in the tender proposal.

Signed _____

Tenderer _____